



## Uttar Pradesh Textile Technology Institute, Kanpur, 11/208, Souterganj, Kanpur

## **ENQUIRY LETTER**

Package Code: TEQIP-III/2019/UP/upti/174

Current Date: 26-Jul-2019

Package Name: Scanning Electron Microscope

Method: Shopping Goods

Sub: INVITATION LETTER FOR Scanning Electron Microscope

Dear Sir,

 You are invited to submit your most competitive quotation for the following goods with item wise detailed specifications given at Annexure I,

Sr. No	Item Name	Quantity	Place of Delivery	Installation Requirement (if any)
1	Scanning Electron Microscope	1	UPTTI, Kanpur	Yes

- 2. Government of India has received a credit from the International Development Association (IDA) towards the cost of the Technical Education Quality Improvement Programme [TEQIP]-Phase III Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for quotations is issued.
- 3. Quotation
  - 3.1 The contract shall be for the full quantity as described above.
  - 3.2 Corrections, if any, shall be made by crossing out, initialling, dating and re writing.
  - 3.3 All duties and other levies payable by the supplier under the contract shall be included in the unit Price.
  - 3.4 Applicable taxes shall be quoted separately for all items.
  - The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
  - 3.6 The Prices should be quoted in Indian Rupees only.
- 4. Each bidder shall submit only one quotation.
- 5. Quotation shall remain valid for a period not less than 60 days after the last date of quotation submission.
- 6. Evaluation of Quotations: The Purchaser will evaluate and compare the quotations determined to be Substantially responsive i.e. which

- 6.1 are properly signed; and
- 6.2 Confirm to the terms and conditions, and specifications.
- The Quotations would be evaluated for all items together.
- 8. Award of contract The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.
  - Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of Contract.
  - 8.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be Incorporated in the purchase order.
- Payment shall be made in Indian Rupees as follows:
   Bidder must quote Price in INR F.O.R at Store UPTTI, Kanpur.

Payment Description	Expected Delivery Period (in Days)	Payment Percentage
Satisfactory Delivery & Installation	30	90
Satisfactory Acceptance	30	10

10.	Liquidated Damages will be applied as per the below:
	Liquidated Damages Per Day Min %:N/A
	Liquidated Damages Max %:N/A
11.	All supplied items are under warranty of 12 months from the date of successful acceptance of
	items and AMC/Others is 0.
12.	You are requested to provide your offer latest by 16:00 hours on 13-Aug-2019.
13.	Detailed specifications of the items are at Annexure I.
14.	Training Clause (if any) yes
15.	Testing/Installation Clause (if any) yes
16.	Performance Security shall be applicable: 5%( Security must be in the form DD in favour of Director, UPTTI, Kanpur, payable at Kanpur. Performance security will be submitted by lowest Qualified bidder at the time of installation and will be detained up to the expiry of warranty.)

4

por My Slund

(Authorized Signatory)

U.P. Textile Technology Institute

11/208, Soutergani

Kanpur U.P.-208001

Name & Designation

## Annexure I

Sr. Ite	em Name	Specifications
E	canning lectron icroscope	Magnification: X25 to1,00,000X or better with SE detector Electron source: Precantered tungsten filament or equivalent electron source Vacuum System: turbo molecular pump and backed by rotary pump Sample stage dimension: X = 30 mm or bigger, Y =30 mm or bigger, stage movement motorised Detectors: Secondary Electron Detector. 2. Back Scattered Electron Detector. Offered SEM should also give Composition /Topography /Shadow image Lens Configuration: suitable two stage Condenser lens Image format: At least TIFF, JPG, BMP Image Pixel: 1280 X 1024 or better Auto-features: Auto focus, Auto stigmator, Auto Gun Alignment, Auto Filament, Adjustment, Single Touch Switching from High Vacuum Mode to Low Vacuum Mode Sample Dia and Height: Max sample diameter approx 70 mm or higher, sample height approx 50 mm or higher Resolution: 20 nm or better Warranty: Three years from the date of installation (including all parts except consumables), Tool kit has to be supplied with Tabletop system Sputtering Unit: Suitable sputtering coating/other coating unit for non-conducting samples Computer System: High-end Computer system with high capacity Hard disk and CD/DVD writer drive approx. 23" or above Color touch screen LCD monitor. The PC equipped with licensed windows 10 or latest operating system.

## (In letterhead of the supplier with seal)

Date:

		ost	Total Cost			
		and warranty/ guaranty commitments)				
		local costs incidental to delivery				
		excise duty, packing and forwarding, transportation, insurance, other			Specifications)	
In % In figures (B)	(A)	(Including Ex-Factory price,		_	goods \ (with full	
Sales tax and other taxes payable	Total Price	Quoted Unit rate in Rs.	Unit	Qty.	Description of	SI. No.